



Request for Quotation

245 NE 30 Rd.

6027.1

Great Bend KS, 67530

Date: December 21, 2010

Vendor:	Notes
---------	-------

Quantity	Description	Unit Price	Extension
2 each	<p>Gas sampler, probe-mounted, adjustable pump displacement style with micrometer adjustment displacement and 1 in. npt connection for mounting on 4 in. pipe. Stationary mount. Design pressure of 50 in. wc and operating pressure od 23 in. wc. Product is atmospheric air with 0.0581 lb./ft. 3 density, flow rate is 300 ACFM. Sampler will operate with proportional to flow, straight time-based or time-based with on/off sampling. Sampler to have minimum 6 digit LCD counter with diagnostic test functions, sealed switches, local mounted wiring assembly and microprocessor to be battery pack operated. Sampler container to be standard spun with ¼ in. connections. Size shall be 300 cc. Tag 1-AE-201, 2-AE-201</p> <p>Documentation required with order: Operation and Maintenance Manual Spare Parts List Certificate of Compliance</p>		